



Testing Laboratory
Accreditation
Certificate

Accreditation No. RTL04590

*Chiyoda Technol Corporation,
Radiation Monitoring Center*

*3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311
Japan*

meets the following criteria. On the basis of this, Japan Accreditation Board (JAB) grants accreditation to the said testing laboratory.

Applicable accreditation criteria

: JIS Q 17025:2018 (ISO/IEC 17025:2017)

Scope of accreditation

: **Ionizing Radiation Monitoring**

(As described in the appendix)

Premises covered by accreditation

: As described in the appendix.

Expiry date of accreditation

: March 31, 2027

Revised
Renewed
Initial accreditation

August 2, 2024
April 1, 2023
March 19, 2019

Y. Miki
Y. Miki, President

Japan Accreditation Board

Accreditation Certificate

Appendix

(Page 1/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

1) Premises on which testing activities are performed

Name of Premises	Radiation Monitoring Center, Customer Service Section	
Address of Premises	Postal code	311-1311
	Address	3522 Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki
Testing service at permanent facilities or on site testing service	<input checked="" type="checkbox"/> Testing service at permanent facilities <input type="checkbox"/> On site testing service	

Scope of Accreditation

FIELD	M33 Ionizing Radiation Monitoring
CODE AND NAME OF CLASSIFICATION (1)	M33.1 Personal dose measurement
CODE AND NAME OF CLASSIFICATION (2)	M33.1.1 Whole body dosimeter

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
FX	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 80 keV	0.1 mSv to 2 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 80 keV	0.1 mSv to 2 Sv
FS	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FV	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv

Accreditation Certificate Appendix

(Page 2/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
FJ	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FR	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FT	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FK	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FN	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv


Accreditation Certificate

Appendix

(Page 3/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
FL	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
NS	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
	M33.1.1.3 Neutron	M33.1.1.3.1 $H_p(10)$	0.025 eV to 15 MeV	0.1 mSv to 60 mSv
NR	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
	M33.1.1.3 Neutron	M33.1.1.3.1 $H_p(10)$	0.025 eV to 15 MeV	0.1 mSv to 60 mSv
NT	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
	M33.1.1.3 Neutron	M33.1.1.3.1 $H_p(10)$	0.025 eV to 15 MeV	0.1 mSv to 60 mSv




Accreditation Certificate Appendix

(Page 4/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
NK	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
	M33.1.1.3 Neutron	M33.1.1.3.1 $H_p(10)$	0.025 eV to 15 MeV	0.1 mSv to 60 mSv
NN	M33.1.1.3 Neutron	M33.1.1.3.1 $H_p(10)$	0.025 eV to 15 MeV	0.1 mSv to 60 mSv
FD	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	80 keV to 6.4 MeV	0.1 mSv to 10 Sv



Accreditation Certificate

Appendix

(Page 5/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

Scope of Accreditation

FIELD	M33 Ionizing Radiation Monitoring
CODE AND NAME OF CLASSIFICATION (1)	M33.1 Personal dose measurement
CODE AND NAME OF CLASSIFICATION (2)	M33.1.2 Extremity dosimeter

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
FW	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	24 keV to 1.25 MeV	0.1 mSv to 10 Sv
	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FU	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	24 keV to 1.25 MeV	0.1 mSv to 10 Sv
	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
JQ	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	16 keV to 1.25 MeV	0.1 mSv to 1 Sv
JR	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	16 keV to 1.25 MeV	0.1 mSv to 1 Sv
JS	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.8 MeV	0.2 mSv to 1 Sv
JT	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.8 MeV	0.2 mSv to 1 Sv
TS	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	16 keV to 1.25 MeV	0.2 mSv to 2 Sv
	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.8 MeV	0.2 mSv to 2 Sv
TH	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	83 keV to 1.25 MeV	0.2 mSv to 2 Sv
	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.8 MeV	0.2 mSv to 2 Sv



Accreditation Certificate Appendix

(Page 6/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

Scope of Accreditation

FIELD	M33 Ionizing Radiation Monitoring
CODE AND NAME OF CLASSIFICATION (1)	M33.1 Personal dose measurement
CODE AND NAME OF CLASSIFICATION (2)	M33.1.3 Eye-lens dosimeter

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
LA	M33.1.3.1 Photon radiation	M33.1.3.1.1 $H_p(3)$	24 keV to 1.25 MeV	0.1 mSv to 1 Sv
	M33.1.3.2 Beta radiation	M33.1.3.2.1 $H_p(3)$	0.8 MeV	0.1 mSv to 1 Sv
LH	M33.1.3.1 Photon radiation	M33.1.3.1.1 $H_p(3)$	24 keV to 1.25 MeV	0.1 mSv to 1 Sv
	M33.1.3.2 Beta radiation	M33.1.3.2.1 $H_p(3)$	0.8 MeV	0.1 mSv to 1 Sv
SS	M33.1.3.1 Photon radiation	M33.1.3.1.1 $H_p(3)$	24 keV to 1.25 MeV	0.1 mSv to 10 Sv
	M33.1.3.2 Beta radiation	M33.1.3.2.1 $H_p(3)$	0.8 MeV	0.1 mSv to 10 Sv

Scope of Accreditation

FIELD	M33 Ionizing Radiation Monitoring
CODE AND NAME OF CLASSIFICATION (1)	M33.2 Personal dose assessment

CODE AND NAME OF CLASSIFICATION (2)	CODE AND NAME OF CLASSIFICATION (3)	TYPE OF IONIZED RADIATION	REMARKS
M33.2.1 Uniform exposure to the body trunk	M33.2.1.1 Effective dose	Photon radiation, Neutron	
	M33.2.1.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.1.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	

Accreditation Certificate Appendix

(Page 7/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

CODE AND NAME OF CLASSIFICATION (2)	CODE AND NAME OF CLASSIFICATION (3)	TYPE OF IONIZED RADIATION	REMARKS
	M33.2.1.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.2 Non-uniform exposure to the body trunk	M33.2.2.1 Effective dose	Photon radiation, Neutron	
	M33.2.2.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.2.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.2.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.3 Uniform exposure to the body trunk and extremity exposure	M33.2.3.1 Effective dose	Photon radiation, Neutron	
	M33.2.3.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.3.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.3.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.4 Non-uniform exposure to the body trunk and extremity exposure	M33.2.4.1 Effective dose	Photon radiation, Neutron	
	M33.2.4.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.4.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.4.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.5 Uniform exposure to the body trunk and eye lens exposure	M33.2.5.1 Effective dose	Photon radiation, Neutron	
	M33.2.5.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	

Accreditation Certificate Appendix

(Page 8/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

CODE AND NAME OF CLASSIFICATION (2)	CODE AND NAME OF CLASSIFICATION (3)	TYPE OF IONIZED RADIATION	REMARKS
	M33.2.5.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.5.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.6 Non-uniform exposure to the body trunk and eye lens exposure	M33.2.6.1 Effective dose	Photon radiation, Neutron	
	M33.2.6.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.6.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.6.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.7 Uniform exposure to the body trunk and extremity exposure and eye lens exposure	M33.2.7.1 Effective dose	Photon radiation, Neutron	
	M33.2.7.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.7.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.7.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.8 Non-uniform exposure to the body trunk and extremity exposure and eye lens exposure	M33.2.8.1 Effective dose	Photon radiation, Neutron	
	M33.2.8.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.8.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.8.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	



Accreditation Certificate Appendix

(Page 9/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan


1) Premises on which testing activities are performed

Name of Premises	Radiation Monitoring Center Aomori	
Address of Premises	Postal code	039-3212
	Address	1-86, Iyasakiadaira, Obuchi, Rokkasyo, Aomori
Testing service at permanent facilities or on site testing service	<input checked="" type="checkbox"/> Testing service at permanent facilities <input type="checkbox"/> On site testing service	

Scope of Accreditation

FIELD	M33 Ionizing Radiation Monitoring		
CODE AND NAME OF CLASSIFICATION (1)	M33.1 Personal dose measurement		
CODE AND NAME OF CLASSIFICATION (2)	M33.1.1 Whole body dosimeter		

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
FJ	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FR	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FN	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv



JAC-MRA

Accreditation No.

RTL04590

Accreditation Certificate Appendix

(Page 10/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
FL	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
NS	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
	M33.1.1.3 Neutron	M33.1.1.3.1 $H_p(10)$	0.025 eV to 15 MeV	0.1 mSv to 60 mSv
NR	M33.1.1.1 Photon radiation	M33.1.1.1.1 $H_p(10)$	16 keV to 6.4 MeV	0.1 mSv to 10 Sv
		M33.1.1.1.2 $H_p(0.07)$	12 keV to 6.4 MeV	0.1 mSv to 10 Sv
	M33.1.1.2 Beta radiation	M33.1.1.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
	M33.1.1.3 Neutron	M33.1.1.3.1 $H_p(10)$	0.025 eV to 15 MeV	0.1 mSv to 60 mSv

Accreditation Certificate

Appendix

(Page 11/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

Scope of Accreditation


FIELD	M33 Ionizing Radiation Monitoring			
CODE AND NAME OF CLASSIFICATION (1)	M33.1 Personal dose measurement			
CODE AND NAME OF CLASSIFICATION (2)	M33.1.2 Extremity dosimeter			

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
FW	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	24 keV to 1.25 MeV	0.1 mSv to 10 Sv
	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv
FU	M33.1.2.1 Photon radiation	M33.1.2.1.1 $H_p(0.07)$	24 keV to 1.25 MeV	0.1 mSv to 10 Sv
	M33.1.2.2 Beta radiation	M33.1.2.2.1 $H_p(0.07)$	0.2 MeV to 0.8 MeV	0.1 mSv to 10 Sv

Scope of Accreditation

FIELD	M33 Ionizing Radiation Monitoring			
CODE AND NAME OF CLASSIFICATION (1)	M33.1 Personal dose measurement			
CODE AND NAME OF CLASSIFICATION (2)	M33.1.3 Eye lens dosimeter			

TYPE OF DOSEMETER	CODE AND NAME OF CLASSIFICATION (3)	CODE AND NAME OF CLASSIFICATION (4)	ENERGY RANGE OF IONIZED RADIATION	RANGE OF DOSE
SS	M33.1.3.1 Photon radiation	M33.1.3.1.1 $H_p(3)$	24 keV to 1.25 MeV	0.1 mSv to 10 Sv
	M33.1.3.2 Beta radiation	M33.1.3.2.1 $H_p(3)$	0.8 MeV	0.1 mSv to 10 Sv



Accreditation Certificate Appendix

(Page 12/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

Scope of Accreditation

FIELD	M33 Ionizing Radiation Monitoring
CODE AND NAME OF CLASSIFICATION ⁽¹⁾	M33.2 Personal dose assessment

CODE AND NAME OF CLASSIFICATION ⁽²⁾	CODE AND NAME OF CLASSIFICATION ⁽³⁾	TYPE OF IONIZED RADIATION	REMARKS
M33.2.1 Uniform exposure to the body trunk	M33.2.1.1 Effective dose	Photon radiation, Neutron	
	M33.2.1.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.1.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.1.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.2 Non-uniform exposure to the body trunk	M33.2.2.1 Effective dose	Photon radiation, Neutron	
	M33.2.2.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.2.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.2.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.3 Uniform exposure to the body trunk and extremity exposure	M33.2.3.1 Effective dose	Photon radiation, Neutron	
	M33.2.3.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.3.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.3.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	

Accreditation Certificate

Appendix

(Page 13/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

CODE AND NAME OF CLASSIFICATION (2)	CODE AND NAME OF CLASSIFICATION (3)	TYPE OF IONIZED RADIATION	REMARKS
M33.2.4 Non-uniform exposure to the body trunk and extremity exposure	M33.2.4.1 Effective dose	Photon radiation, Neutron	
	M33.2.4.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.4.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.4.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.5 Uniform exposure to the body trunk and eye lens exposure	M33.2.5.1 Effective dose	Photon radiation, Neutron	
	M33.2.5.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.5.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.5.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.6 Non-uniform exposure to the body trunk and eye lens exposure	M33.2.6.1 Effective dose	Photon radiation, Neutron	
	M33.2.6.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.6.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.6.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	

Accreditation Certificate

Appendix

(Page 14/14)

Type of Laboratory	Testing Laboratory
Name of Laboratory	Chiyoda Technol Corporation, Radiation Monitoring Center
Address	3522, Onuki-cho, Oarai, Higashi-Ibaraki, Ibaraki, 311-1311 Japan

CODE AND NAME OF CLASSIFICATION (2)	CODE AND NAME OF CLASSIFICATION (3)	TYPE OF IONIZED RADIATION	REMARKS
M33.2.7 Uniform exposure to the body trunk and extremity exposure and eye lens exposure	M33.2.7.1 Effective dose	Photon radiation, Neutron	
	M33.2.7.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.7.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.7.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	
M33.2.8 Non-uniform exposure to the body trunk and extremity exposure and eye lens exposure	M33.2.8.1 Effective dose	Photon radiation, Neutron	
	M33.2.8.2 Equivalent dose (Skin)	Photon radiation, Beta radiation, Neutron	
	M33.2.8.3 Equivalent dose (Lens of the eye)	Photon radiation, Beta radiation, Neutron	
	M33.2.8.4 Equivalent dose (Female abdomen)	Photon radiation, Neutron	

(Notes on Accreditation Certificate)

The laboratory is only accredited for laboratory activities outlined within the methods listed above. Reference to any other activity within these standards, such as risk management or risk assessment, does not fall within the laboratory's accredited capabilities.

When version information of standards or methods are not identified in the scope, laboratories shall adapt to use the current version of such standards within six months at latest from the issued date of current version.

Notes for EMC test laboratory for FCC

Accreditation does not imply acceptance to the FCC equipment authorization program. Please see the FCC website (<https://apps.fcc.gov/oetcf/eas/>) for a listing of FCC approved laboratories.

Japan Accreditation Board