



JAB



Calibration Laboratory

Accreditation Certificate

Accreditation No. RCL00160

TECHNORENT CO., LTD.
BUSINESS MANAGEMENT HEADQUARTERS
TOKYO TECHNICAL CENTER

1-5-21, Katsushima, Shinagawa-ku, Tokyo, 140-8580 Japan

meets the following criteria. On the basis of this, Japan Accreditation Board (JAB) grants accreditation to the said calibration laboratory.

Applicable accreditation criteria	: JIS Q 17025:2005 (ISO/IEC 17025:2005)
Scope of accreditation	: Electromagnetics(DC/Low Frequency) (As described in the appendix)
Premises covered by accreditation	: As described in the appendix.
Expiry date of accreditation	: June 30, 2021

Revised (9)	December 6, 2018
Renewed (4)	April 17, 2017
Initial accreditation	June 26, 2001

T. Oda, Chairman
Laboratory Accreditation Committee

Y. Iizuka, President
Japan Accreditation Board

Accreditation No.

RCL00160

Accreditation Certificate

Appendix

(Page 1/1)



JAB



Type of Laboratory	Calibration Laboratory
Name of Laboratory	TECHNORENT CO., LTD. BUSINESS MANAGEMENT HEADQUARTERS TOKYO TECHNICAL CENTER
Address	1-5-21, Katsushima, Shinagawa-ku, Tokyo, 140-8580 Japan

1) Premises on which calibration activities are performed

Name of Premises	TECHNORENT CO., LTD. BUSINESS MANAGEMENT HEADQUARTERS TOKYO TECHNICAL CENTER	
Address of Premises	Postal Code	140-8580
	Address	1-5-21, Katsushima, Shinagawa-Ku, Tokyo Japan
Calibration service at permanent facilities or on site calibration service	<input checked="" type="checkbox"/> Calibration service at permanent facilities <input type="checkbox"/> On site calibration service	

Scope of Accreditation

CODE OF CLASSIFICATION, QUANTITY MEASURAND / CALIBRATION ITEM	RANGE OF CALIBRATION	EXPANDED UNCERTAINTY (APPROXIMATELY 95 % COVERAGE PROBABILITY, $k = 2$)	CALIBRATION PROCEDURE, REMARKS
M11.6 DC voltage Generator	100 mV	23 ppm	W 校 C2520 (in-house method) Reference Standard : Agilent 3458A/HFL
	1 V	9 ppm	
	10 V	8 ppm	
	100 V	10 ppm	
	1000 V	23 ppm	
Measuring Instrument	100 mV	15 ppm	W 校 C2520 (in-house method) Reference Standard : FLUKE 5700A
	1 V	9 ppm	
	10 V	9 ppm	
	100 V	10 ppm	
	1000 V	11 ppm	

Japan Accreditation Board