

Calibration Laboratory

Accreditation
Certificate

Accreditation No.RCL00180



JAB



SUNSEA JAPAN CO., LTD.
ENGINEERING CENTER CALIBRATION DEPARTMENT

931-2, Naka, Yaita-shi, Tochigi, 329-2135 Japan

meets the following criteria. On the basis of this, Japan Accreditation Board (JAB) grants accreditation to the said calibration laboratory.

Applicable accreditation criteria	: JIS Q 17025:2005 (ISO/IEC 17025:2005)
Scope of accreditation	: Electromagnetics(DC/Low Frequency) (As described in the appendix)
Premises covered by accreditation	: As described in the appendix.
Expiry date of accreditation	: August 31, 2021

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system.

The management system requirements in ISO/IEC 17025:2005 meet the principles of ISO 9001:2008 and are aligned with its pertinent requirements.

Renewed (4)
Initial accreditation

May 12, 2017
August 24, 2001

T. Oda, Chairman
Laboratory Accreditation Committee

Y. Mizuka, President
Japan Accreditation Board

Accreditation No.

RCL00180

Accreditation Certificate

Appendix

(Page 1/1)



JAB



Type of Laboratory	Calibration Laboratory
Name of Laboratory	SUNSEA JAPAN CO., LTD. ENGINEERING CENTER CALIBRATION DEPARTMENT
Address	931-2, Naka, Yaita-shi, Tochigi, 329-2135 Japan

1) Premises on which calibration activities are performed

Name of Premises	SUNSEA JAPAN CO., LTD. ENGINEERING CENTER CALIBRATION DEPARTMENT	
Address of Premises	Postal Code	329-2135
	Address	931-2, Naka, Yaita-shi, Tochigi, Japan
Calibration service at permanent facilities or on site calibration service	<input checked="" type="checkbox"/> Calibration service at permanent facilities <input type="checkbox"/> On site calibration service	

Scope of Accreditation

CODE OF CLASSIFICATION, QUANTITY MEASURAND / CALIBRATION ITEM	RANGE OF CALIBRATION	EXPANDED UNCERTAINTY (APPROXIMATELY 95 % COVERAGE PROBABILITY, $k = 2$)	CALIBRATION PROCEDURE, REMARKS
M11.6 DC voltage Generator Voltage (+)	100 mV	18 ppm	SKC-001 (in-house method) Reference Standard : HP 3458A DC voltage generator (variable type, resolution : 1 ppm or better)
	1 V	7 ppm	
	10 V	5 ppm	
	100 V	8 ppm	
	1000 V	22 ppm	
Measuring Instrument Voltage (+)	100 mV	14 ppm	SKC-002 (in-house method) Reference Standard : FLUKE5700A
	1 V	8 ppm	
	10 V	8 ppm	
	100 V	9 ppm	
	1000 V	10 ppm	

Japan Accreditation Board