



Calibration Laboratory

Accreditation Certificate

Accreditation No. RCL00180



SUNSEA JAPAN CO., LTD.
ENGINEERING CENTER CALIBRATION DEPARTMENT

931-2, Naka, Yaita-shi, Tochigi, 329-2135 Japan

meets the following criteria. On the basis of this, Japan Accreditation Board (JAB) grants accreditation to the said calibration laboratory.

Applicable accreditation criteria	: JIS Q 17025:2018 (ISO/IEC 17025:2017)
Scope of accreditation	: Electromagnetics(DC/Low Frequency) (As described in the appendix)
Premises covered by accreditation	: As described in the appendix.
Expiry date of accreditation	: August 31, 2025

Revised	July 28, 2021
Renewed	September 1, 2021
Initial accreditation	August 24, 2001

Y. Iizuka, President

Japan Accreditation Board



Accreditation No.

RCL00180

Accreditation Certificate

Appendix

(Page 1/1)

Type of Laboratory	Calibration Laboratory
Name of Laboratory	SUNSEA JAPAN CO., LTD. ENGINEERING CENTER CALIBRATION DEPARTMENT
Address	931-2, Naka, Yaita-shi, Tochigi, 329-2135 Japan

1) Premises on which calibration activities are performed

Name of Premises	SUNSEA JAPAN CO., LTD. ENGINEERING CENTER CALIBRATION DEPARTMENT		
Address of Premises	Postal Code	329-2135	
	Address	931-2, Naka, Yaita-shi, Tochigi Japan	
Calibration service at permanent facilities or on site calibration service	<input checked="" type="checkbox"/> Calibration service at permanent facilities <input type="checkbox"/> On site calibration service		

Scope of Accreditation

CODE OF CLASSIFICATION, QUANTITY MEASURAND / CALIBRATION ITEM	RANGE OF CALIBRATION	EXPANDED UNCERTAINTY ¹⁾	CALIBRATION PROCEDURE, REMARKS
M11.6 DC voltage Generator Voltage (+)	100 mV 1 V 10 V 100 V 1000 V	18 ppm 7 ppm 5 ppm 8 ppm 22 ppm	SKC-001 (in-house method) Reference Standard : HP 3458A DC voltage generator (variable type, resolution : 1 ppm or better)
Measuring Instrument Voltage (+)	100 mV 1 V 10 V 100 V 1000 V	14 ppm 8 ppm 8 ppm 9 ppm 10 ppm	SKC-002 (in-house method) Reference Standard : FLUKE5700A
¹⁾ Information on the coverage factor	<input checked="" type="checkbox"/> $k=2$; level of confidence of approximately 95 % <input type="checkbox"/> coverage factor obtained from the effective degrees of freedom that defines a level of confidence of 95 %, based on the t -distribution <input type="checkbox"/> others ()		

Japan Accreditation Board

Issue No. : RCL00180-20210728